

Lyric4

0.4 cm³ coupler data


FRYE CIC coupler






2 cm³ coupler data

ANSI / ASA S3.22-2014 IEC 60118-0 : 2015





General test information

- All data obtained are measured with Phonak Target measurement settings
- The device is operating in linear mode
- The 0.4 cm³ coupler data is used as additional information which is closer to the real application, as the deep insertion leads to a significantly smaller residual volume in front of the ear drum

WARNING: Changes or modifications to the hearing aid that are not explicitly approved by Sonova AG are not permitted. Such changes may damage the ear or the hearing aid.







