Phonak Naída™ P

The second secon


Phonak Naída P-UP (P90/P70/P50/P30/Trial)


Standard Earhook HE11 680 2 cm³ coupler data

ANSI / ASA S3.22-2014 (R2020) IEC 60118-0: 2022


Standard Earhook HE11 2 cm³ coupler data


ANSI / ASA S3.22-2014 (R2020) IEC 60118-0: 2022














General test information

- Specific measurement settings are used. RTS adjustment with volume control
- The device is operating in linear mode
- Low-level expansion is active
- All data obtained are measured with Phonak Target measurement settings
- \bullet The latency of the audio signal determined according an internal standard is 6.2 ms

Warnings

⚠

This hearing instrument has an output sound pressure level that can exceed 132 dB SPL. Special care should be taken when fitting this instrument as there is a risk of impairing the residual hearing of the user.



Changes or modifications to the hearing aid that are not explicitly approved by the manufacturer are not permitted. Such changes may damage the ear or the hearing aid.

* Battery performance depends on active features, the use of wireless accessories, hearing loss, battery age, sound environment and earpiece. Note that for non-rechargeable ZnAir batteries operating time might vary depending on the battery model



