

Phonak Audéo P-R Fit (P90/P70/Trial)

M Receiver 2 cm³ coupler data


ANSI / ASA S3.22-2014 IEC 60118-0 : 2015






P Receiver 2 cm³ coupler data

ANSI / ASA S3.22-2014 IEC 60118-0 : 2015





General test information

- Specific measurement settings are used. RTS adjustment with volume control
- The device is operating in linear mode
- Low-level expansion is active
- All data obtained are measured with Phonak Target measurement settings



Changes or modifications to the hearing aid that are not explicitly approved by the manufacturer are not permitted. Such changes may damage the ear or the hearing aid.



^{*} Expected operating time of the rechargeable battery depends on active features, the use of wireless accessories, hearing loss, battery age and sound environment.