

Phonak Target 11

Phonak Target/ALPS Fitting Guide

This guide contains details on using Phonak Target/ALPS available in Phonak Target fitting software. Phonak Target/ALPS allows you to enter data via Phonak Target directly into ALPS. This one step process avoids duplicate entry and allows you to manage all relevant Lyric fitting and subscription information in an easy way.

To use Phonak Target/ALPS effectively, make sure your computer has an internet connection.

For every Lyric fitting, Phonak Target will automatically connect you to ALPS. A wizard will guide you through all the necessary steps to complete a successful Lyric fitting. Please note that the Lyric device activation code must be entered the first time Phonak Target is launched.

More information is also be available under Web Help.


Content

Create a new client in Phonak Target	2
Access an existing client in your ALPS account	6
Device exchange (including ear canal observation)	7
Tourist case: Device exchange for a tourist client	10
Tourist case: Your client has been refit by a different Authorized Lyric Partner	11
Renew a subscription	12
Cancellation of an active subscription	13
Offline fitting and synchronization	14
ALPS home and reports	16
Enhanced log files for support	20
Information and description of symbols and System Requirements	20

Create a new client in Phonak Target

To create a new client in Phonak Target:

1. Click on [Instruments] and select Lyric devices in [Hearing aids]
2. Click on [Lyric configuration] and click on [Select device] to open the Phonak Target/ALPS wizard.



The ALPS button indicates your current connection status:


- Green - connected
- Black – disconnected




To connect into ALPS, enter your user name and password to log into your Lyric account.

We strongly recommend that you select the [Automatic login] function. This time-saving feature enables you to remain logged into ALPS without the need to reenter your credentials. That only applies if you're not working on a shared workstation.


In case you have forgotten your password, click on the link [forgot password] and change it in www.phonakpro.com directly.



You will be asked to [Add new client to your ALPS account]. Your account information is shown on this screen, that is very helpful for providers working with multiple logins.



In case you don't intend to login with the last login shown, choose [Switch login] under the ALPS button and login with another username.



You can now see the client’s first and last name, date of birth and gender as entered in Phonak Target.

The additional information (date of birth and gender) is optional. You can uncheck [Submit additional information] if you do not wish to transmit this data.

For new subscriptions you are then asked to confirm that the client understands and gave consent that his/her data will be transferred to Sonova for processing. If the box is not ticked, the fitting cannot continue.

To select the appropriate Lyric device, choose the measured seal size and pick the first device in the list as it is the closest to expire. Devices about to expire within 30 days are marked red.

Please note that expired devices can no longer be fit. To return an expired device, use the [device return process] in ALPS, select the serial number and the reason for device return as “expired” and send it back to Phonak.

Select the measured insertion depth.

Choose the subscription duration you want to create for your client. Make sure to click the check box, in case the customer is using Lyric for the treatment of tinnitus.

Lyric

Enter client details

First name

Last name

Date of birth

Gender

☒ Submit additional information

☐ I hereby confirm that I fully informed my client that his personal data will be transferred to the hearing aid manufacturer for processing purpose.
My client consented to such transfer.

Create Lyric ID Cancel

Lyric

Select Lyric device from consignment stock

Former device seal size: XXL

Serial#

Insertion date

Serial#	Model	Seal size	XXL	Insert by
2307NYSEH	Lyric4	XXL		30.03.2023
2307NYSEK	Lyric4	XXL		05.04.2023
2307NYSEL	Lyric4	XXL		05.04.2023
2307NYSEJ	Lyric4	XXL		09.04.2023
2307NYSEF2	Lyric4	XXL		13.04.2023
2307NYSEP	Lyric4	XXL		17.02.2024
2307NYSEW	Lyric4	XXL		17.02.2024
2307NYSEV	Lyric4	XXL		17.02.2024
2307NYSEU	Lyric4	XXL		17.02.2024

Select Cancel

Lyric

Select measured insertion depth

Initial device insertion for this ear.

--select--

3 mm

4 mm

5 mm

6 mm

7 mm

8 mm

9 mm

10 mm

11 mm

12 mm

13 mm

14 mm

15 mm

Continue Cancel

Lyric

Select a subscription

Subscription type


Payment type

Start date


☐ Lyric is used for tinnitus treatment

Continue Cancel


Phonak Target confirms the device selection. You can get [More information] about the subscription in this window or choose [Close]. For binaural fittings, you are asked to [Proceed with other side] which leads you directly to selecting another device from your consignment stock.




Below the start and end date of the subscription and actual days remaining to the trial end date are shown. It is now possible to extend a trial once for an amount of 10 days. The new end date will then be shown in Target/ALPS.




In case a wrong serial number has been chosen, it can be returned to the stock by clicking “return to stock”.




You have the possibility to choose between Gain (Real ear and Coupler 0.4cc) and Output display (SPL real ear and coupler 0.4cc) when fitting Lyric. Click on the [Output] menu to select your desired output curve display. When selecting the output curve display, you may also choose to show or hide the hearing level and UCL.




In case the achieved insertion depth varies from the measured one, a measured and an achieved insertion depth can be entered in [Fitting]. A warning will be shown if the achieved insertion depth is higher than the measured one.




By clicking [Save to database] at the end of the session, your fitting data will be synchronized with ALPS.




In case you need to have more notes stored around the patient, fill them in on the new notes section.



A new note can be added and Target allows a “remind me again in next session” function.



All note are then visible under [Client, Notes].




The following reports can be printed from the Print menu:

For Professional:

- Fitting Report
- Lyric Datasheet


For Client:

- Hearing Aid Instructions (SoundLync)
- Communication tips
- Diary of observation
- Understanding hearing loss



The curves reports can be printed as follows:

Choose [Setup], [Reports], [Curve type] and choose between Gain or Output. The selected curve type will then be visible on the curve report.



Access an existing client in your ALPS account

If the client exists in ALPS but not in Target (or in case of a first access within a satellite office), you can [Access existing client in your ALPS account].

First, create a new client in Noah and enter the client’s data. This will be transferred into Phonak Target. Proceed until you reach the screen where you are asked to add or access an existing client.

Enter the correct Lyric ID and/or the first and last names as stored in ALPS.

To retrieve the Lyric ID you can open the ALPS portal in a separate browser or get it from the printed Lyric ID card. Hit search to find P# (Lyric ID).

The Lyric ID has to be entered only if you are connecting an existing client for the first time. When opening the session again, you will automatically be connected to the ALPS data.

By selecting the search button the P# will populate, if the P# does not populate, this means the patient is not an existing patient or the name does not match the ALPS information.

If the data of the client was found, the client subscription information screen will show you all the relevant information.

Lyric

Client has no Lyric ID assigned in Phonak Target

Account 0020035121 LYRIC PORTAL - TARGET TEAM

☐

 Add new client to your ALPS account

☒

 Access existing client in your ALPS account

☐

 Access existing client in a different ALPS account (Tourist case)

Continue

Cancel

Lyric

Lyric client

First name

John

Last name

Doe

Lyric ID

P#####

Date of birth

-

Gender

-

Search

Continue

Cancel

Lyric

Client subscription information

First name

John

Last name

Doe

Lyric ID

P000087661

Information	Ⓜ Right	Ⓛ Left
Subscription	Trial ✓ (20 days left)	No subscription
Start date	06.06.2014	
End date	05.06.2015	
Remaining devices	11	
Device	Lyric3	
Serial number	1406NY0JT	

Continue

Cancel

Device exchange (including ear canal observation)

To exchange a Lyric device, please log into Phonak Target/ALPS, go to [Instruments], [Lyric configuration] and click [Exchange/remove device].

Subscription

Status: Trial (27 days left)

Start date: 07.11.2017

End date: 07.12.2017

Trial end date: 07.12.2017

Billing date: 07.12.2017

Remaining devices: 1

Lyric3

Serial number: 1733NY3HT (Code: E325)

Seal size: S

Insertion date: 09.11.2017

Insertion depth: Measured 10 mm

Exchange/remove device

Select if the provider or the patient has removed the device and specify the removal date if not today. Choose the corresponding removal reason. The removal reasons are separated into four categories (Discomfort, Device related, Fitting related, Proactive removal, Self-replacement and Cancelled Subscription).

Remove Lyric device

Serial number: 2125NY1A9

Removed by: ☒ Provider ☐ Patient self-removed

Removal date: ☒ Today ☐ Previously

Removal reason: --select--

- Discomfort
- Device related
- Fitting related
- Proactive removal
- Self-replacement
- Cancelled subscription

Remove Cancel

Hovering over individual removal reasons activates the “tooltips”, which gives you more information to help you understand which reason to select.

Remove Lyric device

Serial number: 2125NY1A9

Removed by: ☒ Provider ☐ Patient self-removed

Removal date: ☒ Today ☐ Previously

Removal reason: --select--

- Discomfort
- Device related
 - Device became weak
 - No/intermittent response to sound
 - Device performance decreases over time
 - Sound artifacts
- Fitting related
- Proactive removal
- Self-replacement
- Cancelled subscription

Remove Cancel

Phonak Target will ask you for the status of the ear canal. You can choose between: [OK for immediate refit], [OK for refit after rest] and [Requires medical referral]. When choosing [OK for immediate refit], the device exchange can happen immediately. In case the ear canal was irritated and needed some rest or a medical referral, the system will inform you accordingly.*

The comments for removal reason and ear canal observation are both optional.

In case [OK for refit after rest] or [Requires medical referral] is chosen, additional details about the state of the ear canal must be chosen. Additionally, if “other reason” is chosen, a comment to give more explanation is required.

To remove a second device, click on [Proceed with other side]. There are some reasons (scheduled removal, near end of device life and clinical proactive removal) that will give the second side the same reason as a default but it can be changed if the removal reason is different.

*If you have chosen [OK for refit after rest] or [Requires medical referral], Phonak Target will inform you on the main screen.

Lyric

Remove Lyric device

Serial number

1633H00KR

Removed by

☐ Provider

☒ Patient self-removed

Removal date

☒ Today

☐ Previously

Removal reason

Discomfort

Additional details (optional)...

Ear canal observation

OK for immediate refit

Additional ear canal observation (optional)...

Remove

Cancel

Lyric

Remove Lyric device

Serial number

1849NY3G3

Removed by

☒ Provider

☐ Patient self-removed

Removal date

☒ Today

☐ Previously

Removal reason

Discomfort

Additional details (optional)...

Ear canal observation

Requires medical referral

Additional details

--select--

Redness of tissue

Excess fluid collection on tissue

Swelling of clotted blood below tissue (bruise)

Sore or ulceration of tissue

Medial bulge/growth

Blood/Bleeding

Cerumen management

Other reason

Remove

Cancel

Lyric

Successfully removed

Device successfully removed.

Print return form

Select new device

Close

Select device

OK for refit after rest (13.11.2017)

Redness of tissue - Mild

Before you can exchange a device, it is mandatory to confirm that the [Ear is ok for immediate fit].

Lyric

Remove Lyric device

Serial number: 1736NVHCG

Removed by: ☒ Provider ☐ Patient self-removed

Removal date: ☒ Today ☐ Previously

Removal reason: Discomfort

Additional details (optional)...

Ear canal observation: OK for immediate refit

Additional ear canal observation (optional)...

Remove Cancel

After confirming the ear is OK for immediate fit, choose another Lyric device to exchange.

Lyric

Select Lyric device from consignment stock

Former device seal size: M

Serial... Insert date: 16.06.2014

Serial#	Model	Seal size	Insert by
1406NY0JV	Lyric3	M	08.05.2015
1406NY0JN	Lyric3	M	08.05.2015
1406NY0K0	Lyric3	M	08.05.2015
1406NY0IX	Lyric3	M	08.05.2015
1406NY0JJ	Lyric3	M	08.05.2015
1406NY0JH	Lyric3	M	08.05.2015
1406NY0JM	Lyric3	M	08.05.2015
1406NY0JK	Lyric3	M	08.05.2015

Select Cancel

The comments you entered are saved under [Client, Notes].

Lyric

Fitting

Notes

Enter new session note...

16.06.2014 13:57: ECO Right: OK for refit after rest.
ECO Right: OK for refit after rest.
Ear canal observation after device removal.
Ear canal is a bit irritated.

16.06.2014 13:56: ECO Left: OK for immediate refit.

16.06.2014 13:55: ECO Right: OK for immediate refit.

Please note that if the device has been worn longer than 30 days, the [Easy Refit] is enabled and will show default values:

- Removed by provider
- Removed today
- Removal reason: scheduled removal
- Ear canal observation: "ok for immediate refit"

Click [Remove] and then choose a new device accordingly.
In case the criteria doesn't match the default value, please change it and continue.

Lyric

Remove Lyric device

Serial number: 1836NV212

Removed by: ☒ Provider ☐ Patient self-removed

Removal date: ☒ Today ☐ Previously

Removal reason: Scheduled removal

Additional details (optional)...

Ear canal observation: OK for immediate refit

Additional ear canal observation (optional)...


Remove Cancel

Tourist case: Device exchange for a tourist client


If you need to exchange a Lyric device for a client who was originally fit by another Authorized Lyric Partner, you need to get access to the fitting data of that client's ALPS fitting data. To do so you need the client's individual Lyric ID.

To start the tourist case, proceed as you would with a new client. Log into your ALPS account and select [Access existing client in a different ALPS account] (Tourist case)].

Enter the client's Lyric ID, and check if the name is displayed correctly (as printed on the Lyric ID card). Proceed with the exchange of the device, and adjust the fitting parameters if necessary. If the patient doesn't know their Lyric ID, it can be obtained by calling Lyric customer service.



The screenshot shows the 'Lyric client' entry form. It has a green header with the 'Lyric' logo. Below the header, the form fields are: 'First name' (John), 'Last name' (Doe), 'Lyric ID' (P00008766), 'Date of birth' (-), and 'Gender' (-). There is an information icon (i) below the fields. At the bottom, there is a message: 'Please contact Lyric customer support if the client's Lyric ID is unknown.' and two buttons: 'Continue' and 'Cancel'.



The screenshot shows the 'Lyric' client selection screen. It has a green header with the 'Lyric' logo. Below the header, the text reads: 'Client has no Lyric ID assigned in Phonak Target'. Underneath, it says 'Account 0020035121 LYRIC PORTAL - TARGET TEAM'. There are three radio button options: 'Add new client to your ALPS account', 'Access existing client in your ALPS account', and 'Access existing client in a different ALPS account (Tourist case)'. The third option is selected. Below the options, there is an information icon (i) and a message: 'Session data in Phonak Target will be updated from ALPS.' At the bottom, there are two buttons: 'Continue' and 'Cancel'.

Tourist case: Your client has been refit by a different Authorized Lyric Partner

If your client has been refit by a different Lyric Provider, you will be alerted that there is a mismatch between ALPS (tourist fitter) and Phonak Target (your fitting). This indicates that the serial number has been changed.

By clicking [Continue] the system will synchronize the newest serial number.

Lyric

❗ Data conflict between ALPS and Phonak Target

First name John
Last name Doe
Lyric ID P000087661

Device data in ALPS and Phonak Target do not match:

Information

❗ Right

⚙ Left

Subscription	Trial (20 days left)	Trial (30 days left)
ALPS	1406NY0YG (Tourist exchange)	No serial number
Target	No serial number	No serial number

Action	❗ Serial number will be updated in Phonak Target	No action required
--------	--	--------------------

ⓘ Press "Cancel" to resolve the conflicts manually.

Continue

Cancel

If the tourist Lyric fitter also changed the fitting parameters, a fitting parameter mismatch will be shown. You can solve the conflict by choosing the preferred option.

Please note that the ALPS data is the most up to date information from the tourist fitter and Phonak Target data is the data you entered yourself before the tourist fitting.

Lyric

❗ Fitting parameter mismatch!

Fitting parameters stored in ALPS and Phonak Target do not match.

❗ Right

⚙ Left

❗ Resolve conflict

☐ Use Phonak Target data
Last updated 16.06.2014 13:57

☒ Use ALPS data (newer)
Last updated 16.06.2014 14:05
Updated by tourist fitter.

Fitting parameters are identical

ⓘ Clicking "Continue" will resolve the conflict.
Press "Cancel" to resolve the conflict manually.

Continue

Cancel


Renew a subscription

As soon as a subscription is active (no longer in trial) it can be renewed by clicking on [Renew subscription].

When the subscription is three months from expiration, an orange [Renew subscription] button will appear.

Define which type of subscription you want to create in ALPS for your client and follow the instruction wizard.

After renewing a cancelled subscription (trial or active), a new trial will only be permitted after 60 days. The renewed subscription will automatically be an active subscription after renewing within 60 days of the cancellation.



The screenshot shows a software interface with a green header bar containing 'Hearing aids' and 'Lyric configuration'. Below this is a 'Subscription' window. The window displays the following information:



- Status: Active (61 days left)
- Start date: 01.08.2016
- End date: 01.08.2016
- Trial end date: 01.08.2016
- Remaining devices: 7 (in subscription)

On the right side of the window, there are two buttons: 'Cancel subscription' and 'Renew subscription'. The 'Renew subscription' button is highlighted with an orange border. At the bottom right, there is a dropdown menu labeled 'ALPS links'.

Cancellation of an active subscription

If there is a need to cancel an active subscription, you can do so by clicking [Cancel subscription] on the Lyric configuration screen.

The system will make you aware that you are about to cancel an active subscription. To proceed, choose the corresponding cancellation reason and click Continue. The cancellation request will be sent to Phonak.



Offline fitting and synchronization

There might be occasions where you have a client set up via Phonak Target/ALPS, however, due to various reasons (such as temporarily no internet connection or computer problems) you will have one or more sessions where you have to work offline. This means that the new data you have in Phonak Target is not synchronized with ALPS.

Removing a device in offline mode works the same way as in removing a device in online mode.

You are now asked to select a new device.

Lyric configuration

Ⓡ

Remove Lyric device

Serial number

1633H00L5

Removed by

☒ Provider

☐ Patient self-removed

Removal date

☒ Today

☐ Previously

Removal reason

Discomfort

Additional details (optional)...

Ear canal observation

OK for immediate refit

Additional ear canal observation (optional)...

ⓘ

Not connected to ALPS. All changes will be synchronized later.

Remove

Cancel

Lyric configuration

✓

Successfully removed

Ⓡ

Device successfully removed.

Select new device

Close

Enter the serial number as well as the code that is shown on the Lyric package. Enter the measured insertion depth. If needed, the fitting parameters can be changed.

Lyric configuration

Ⓡ

 Enter device information

Serial number

 1633H00L8

Code

 E396

Model

 Lyric3

Seal size

 XL

ⓘ

 Not connected to ALPS. All changes will be synchronized later.

Select

Cancel

Once the internet connection problem is resolved and you are able to use Phonak Target/ALPS again, you will be notified by an orange exclamation point [Fitting not saved to ALPS] message for each patient that was fit offline.

Offline, Sync

Lyric ID P000149579

Session 11 26.10.2016 10:17

Ⓡ

 No audiogram

Ⓛ

 No audiogram

Lyric

SN: 1633H00L8

No hearing instrument

ⓘ

 Fitting not saved to ALPS

As soon as Phonak Target has started connecting into ALPS again, it will automatically synchronize the offline sessions.

Lyric

ⓘ

 Synchronize offline changes

Change	Serial number	Date	Reason	Sync status
Ⓡ Removal	1733NV3HJ	13.11.2017	Discomfort	⌚ Pending
Ⓡ Insertion	1733NV2YW	13.11.2017	Discomfort	⌚ Pending
Ⓡ Removal	1733NV2YW	13.11.2017	Discomfort	⌚ Pending
⌚ Synchronizing...				

Sync to ALPS

Cancel

A green check mark will indicate that the synchronization has been successful.

After an offline fitting, logs on your desktop are available.

Lyric

ⓘ

 Synchronize offline changes

Change	Serial number	Date	Reason	Sync status
Ⓡ Removal	1633H00L5	26.10.2016	Discomfort	✅ Synced
Ⓡ Insertion	1633H00L8	26.10.2016		✅ Synced
✅ Successfully synchronized.				

ⓘ

 One or more device changes could not be synchronized to ALPS!
Click "Continue" to view the current device selection.


Continue

Cancel

ALPS home and reports

There are some actions that require a direct link to ALPS:


- Open ALPS
- Client details
- Device details
- Print Lyric ID card
- Reports



Alternatively, ALPS can also be opened via Phonakpro.com.

On the [Home Screen] you have an overview over:

- Inventory Audit
- Trials about to expire within 7 days
- Subscriptions about to expire within 90 days
- Devices to expire within 120 days



- Print Lyric ID
- Subscription of that client
- Renew Subscription(s)
- Remove Devices (needed when cancelling the subscription)
- Client details (such as fitting parameter)

By clicking on the blue subscription number, more details about the subscription are shown:

- Terms and Conditions
- Cancel Subscription
- Device History

When clicking on the blue device number, this information is available:

- Device State
- Device Code, Warranty & Use By Date
- Subscription/Insertion
- Removal and Return Information

The [Stock] overview shows all devices in the respective account, sorted by device size. That list can also be exported to Excel.

Phonak Target 11 | Phonak Target/ALPS 17

Under [Subscriptions], all subscriptions with all states can be found. When working in a satellite office that is connected in the system, then all clients are visible but separated by account number. That list can be exported to Excel.

Phonak account number	Account Name	Subscription no.	Split ID	First name	Last name	Start date	End date	Side	State	Subscription type	Payment model
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront
2132NY7JY	Lyric Softwax R1 (75)	2132NY7JY		Lyric	Softwax	10/05/2021	10/05/2021	R	True	1 yr subscription	Nearly upfront

The [Device Return] section allows you to return devices to Phonak that are either

- New (overstock)
- Dead on arrival
- Wrong delivery
- Expired

Choose the devices that need to be returned to Phonak. Click on the respective device and choose a return reason. Click select and then print a return form that can be sent with a device to return (if needed).

Serial number	Device size	Side	Short description	Use by
2132NY7JY	S	R/L	Lyric4 (S)	8/6/2022
2132NY7KC	S	R/L	Lyric4 (S)	8/6/2022
2132NY7KA	S	R/L	Lyric4 (S)	8/6/2022
2132NY9WK	XL	R/L	Lyric4 (XL)	8/6/2022
2132NY9WD	XL	R/L	Lyric4 (XL)	8/6/2022
2132NY9WC	XL	R/L	Lyric4 (XL)	8/6/2022
2132NY9WG	XL	R/L	Lyric4 (XL)	8/6/2022
2132NY9WF	XL	R/L	Lyric4 (XL)	8/6/2022
2132NY9WE	XL	R/L	Lyric4 (XL)	8/6/2022
2132NY941	L	R/L	Lyric4 (L)	8/13/2022
2132NY94C	L	R/L	Lyric4 (L)	8/13/2022
2132NY943	L	R/L	Lyric4 (L)	8/13/2022
2132NY948	L	R/L	Lyric4 (L)	8/13/2022
2132NY946	L	R/L	Lyric4 (L)	8/13/2022
2132NY945	L	R/L	Lyric4 (L)	8/13/2022
2132NY8AY	M	R/L	Lyric4 (M)	8/13/2022
2132NY8C0	M	R/L	Lyric4 (M)	8/13/2022
2132NY8C4	M	R/L	Lyric4 (M)	8/13/2022

Under your name, there are the following sections:

- Home
- My Profile
- My Account
- Logout

Target Tester DE

- Home
- My Profile
- My Account
- Logout

[My Profile] does allow you to change your language if needed.

My Settings

Location

Language: English

Time Zone: (GMT+01:00) Central Daylight Time (Automatic Change)

Cancel Save

Under [My Account] you can choose your current location (when working in different locations).

Please choose one account	Phone account number	Account Name	Address	City	Zip code
☐	8333076420	Lyric Testbunde #3	Muensterstr. 1	Stuttgart	70372

In case you'd like to get a weekly e-mail about the trials, subscriptions and stock about to expire, you can fill in up to five e-mail addresses. Also, choose how many months before expiration of the subscriptions you'd like to get a notification. For satellite offices, a report with a full overview over all offices can also be requested by ticking the box.

Notification Settings

Account Name: Lyric Testbunde #3

Notification E-Mail 1:

Notification E-Mail 2:

Notification E-Mail 3:

Notification E-Mail 4:

Notification E-Mail 5:

Expiration Notification Time: 1 month

Notification Report Satellite Offices: ☐

The [Dashboard] gives you an indication about how your business is doing. Explanation is found in more details in the legend within the dashboard.

Trials 64 / New Subscriptions 67 / Renewals 0 / Attivo 3 / Trial Success Rate 4 % / TSR national 25.00 %

1. Trials with starting date Jan 1st - Dec 31st
2. New Subscriptions starting Jan 1st - Dec 31st
3. Renewals with starting date Jan 1st - Dec 31st
4. Total active subscriptions
5. Trial success rate average 12 months
6. Trial success rate average 12 months on a national level

The [Device Analysis] Form can be downloaded.

Forms to Download


The [Search] field does allow you to search data such as subscription, client names, devices etc. directly.

Search...

Enhanced log files for support

If you encounter a problem with Phonak Target, enhanced logging can help our Target software support team to find out the root cause.

Please save the log files on your desktop and/or send them by e-mail to your Lyric contact.



Information and description of symbols and System Requirements

Information and the description of symbols and an overview of system requirements can be found in the Phonak Target User Guide.

This user guide is applicable for Target 11.0 and later subversions of Target 11.0 fitting software. For previous versions of the fitting software user guide contact your local Phonak representative.

Australian Sponsor:
Sonova Australia Pty Ltd
12 Inglewood Place,
Norwest NSW 2153
Australia



Sonova AG • Laubisrütistrasse 28 •
CH-8712 Stäfa • Switzerland



Sonova Deutschland GmbH
Max-Eyth-Str. 20
70736 Fellbach-Oeffingen • Germany

