### Phonak Sky™ M


### Phonak Sky M-SP (M90/M70/M50/M30/Trial)

## Standard and Mini Earhook HE11 2 cm<sup>3</sup> coupler data


ANSI / ASA S3.22-2014 (R2020) IEC 60118-0: 2022


# Standard and Mini Earhook HE11 680 2 cm<sup>3</sup> coupler data


ANSI / ASA S3.22-2014 (R2020) IEC 60118-0: 2022














#### General test information

- $\bullet$  Specific measurement settings are used. RTS adjustment with volume control
- The device is operating in linear mode
- Low-level expansion is active
- All data obtained are measured with Phonak Target measurement settings
- The latency of the audio signal determined according an internal standard is 6.2 ms

#### Warning



This hearing instrument has an output sound pressure level that can exceed 132 dB SPL. Special care should be taken when fitting this instrument as there is a risk of impairing the residual hearing of the user.

A Changes or modifications to the hearing aid that are not explicitly approved by the manufacturer are not permitted. Such changes may damage the ear or the hearing aid.

<sup>\*</sup> Battery performance depends on active features, the use of wireless accessories, hearing loss, battery age, sound environment and earpiece. Note that for non-rechargeable ZnAir batteries operating time might vary depending on the battery model



