

## Declaration of Conformity

ID: 000.000.012.006, V7.0

### Hearing Instrument System Accessories

Roger 10 (02), Roger 11 (02), Roger 13 (02), Roger 14 (02), Roger 15 (02),  
Roger 10 (03), Roger 11 (03), Roger 13 (03), Roger 14 (03), Roger 15 (03),  
Roger 16 (02), Roger 17 (02), Roger 18 (02), Roger 19 (02), Roger X (02),  
Roger 16 (03), Roger 17 (02), Roger 18 (03), Roger 19 (03), Roger X (03)

We, Phonak AG, Laubisrütistrasse 28, 8712 Stäfa, Switzerland, hereby declare under our sole responsibility that the medical device Class I mentioned above, is in conformity with essential requirements of the Medical Device Directive 93/42/ECC (MDD) Annex I and with the Radio Equipment Directive 2014/53/EU and following standards:

Requirement	Standard
HEALTH & SAFETY (Art. 3(1)(a))	IEC 60601-1:2005+Cor.1:2006+Cor.2:2007+A1:2012 EN 60601-1:2006 +Cor.:2010+A1:2013
EMC (Art. 3(1)(b))	ETSI EN 301 489-1 V1.9.2 (2011) ETSI EN 301 489-3 V1.6.1 (2013) EN 60601-1-2:2007 / IEC 60601-1-2:2007 ed.3
RADIO SPECTRUM (Art. 3(2))	ETSI EN 300 440-1 V1.6.1 (2010) ETSI EN 300 440-2 V1.4.1 (2010)

### Additional Information:

This declaration is supported by	Certificate of approval No. 31857 to quality standard ISO 9001:2008 and No.30379 to quality standard ISO 13485:2003 issued by LNE/G-Med
Technical File is held by	Phonak AG, Laubisrütistrasse 28 CH-8712 Stäfa, Switzerland
GMDN Code	57961

Signed on behalf of Phonak AG

Stäfa, June 12, 2017

  
H. Mehl  
Group Vice President  
Operations

Stäfa, June 12, 2017

  
J. Itin  
Director of Corporate TQM & RA